

CONVERSION FACTORS FOR OPERATIONAL QUANTITIES REPORTED BY PACIFIC NORTHWEST
NATIONAL LABORATORY (PNNL)^(a)

For exposures conducted in accordance with ANSI Standard HPS N13.32-1995

Beam Code	Shallow C _x Conversion Factors, rem/R	
	Arm/Leg Phantom (Wrist)	Finger Phantom (Ring)
M30	0.99	0.95
M60	1.04	0.98
M100	1.10	0.99
M150	1.14	1.01
H150	1.14	1.02
¹³⁷ Cs	1.02	0.98
⁶⁰ Co	1.00	0.98

^(a) C_x information taken from Table 3 of ANSI Standard HPS N13.32-1995, Performance Testing of Extremity Dosimeters.